



**Agilent Technologies**

## **Declaration of Conformity** to manufacturing specifications

We herewith inform you that the product  
G1315D with serial number:

**DEAAX02120**

has successfully passed all our production  
quality tests.

During final instrument performance verification  
the following functional characteristics were  
individually tested for conformance with our  
internal specifications:

<b>Leak Sensor Function</b>	<input checked="" type="checkbox"/>
<b>Optical Unit functionality</b>	<input checked="" type="checkbox"/>
<b>Spectral Resolution Test</b>	<input checked="" type="checkbox"/>
<b>Wavelength Accuracy Test</b>	<input checked="" type="checkbox"/>
<b>Straylight Test</b>	<input checked="" type="checkbox"/>
<b>Noise and Drift Test</b>	<input checked="" type="checkbox"/>
<b>IEC 61010-1 Annex F Safety Tests</b>	<input checked="" type="checkbox"/>

September 16, 2011

Signature:   
Test Technician: Roland Becker

**G1315D**



**DEAAX02120**





**Agilent Technologies**

**DECLARATION OF CONFORMITY**  
According to ISO/IEC Guide 22 and CEN/CENELEC EN 45014



**Manufacturer's Name:** Agilent Technologies Singapore (International) Pte. Ltd.  
**Manufacturer's Address:** No. 1 Yishun Ave 7  
**Supplier's Address:** SINGAPORE 768923  
Singapore

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** Diode Array Detector, Multiple Wavelength Detector  
**Model Number:** G1315B, G1315C, G1315D, G1365B, G1365C, G1365D  
**Product Options:** This declaration covers all options of the above products

**complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:**

Low Voltage Directive 2006/95/EC  
EMC Directive 2004/108/EC

**and conforms with the following product standards:**


EMC	Standard	Limit
	IEC 61326-1 :2005 / EN 61326-1:2006 CISPR 11 / EN 55011 IEC 61000-3-2 / EN61000-3-2 IEC 61000-3-3 / EN61000-3-3 IEC 61000-4-2 / EN 61000-4-2 IEC 61000-4-3 / EN 61000-4-3 IEC 61000-4-4 / EN 61000-4-4 IEC 61000-4-5 / EN 61000-4-5 IEC 61000-4-6 / EN 61000-4-6 IEC 61000-4-11 / EN 61000-4-11	Group 1 Class B <sup>[1]</sup>  4kV CD, 4kV AD 1 V/m: 80-1000 MHz, 1.4-2GHz; 1V/m: 2-2.7GHz 0.5 kV signal lines, 1 kV power lines 0.5 kV line-line, 1 kV line-ground 3Vrms, 0.15-80 MHz 0%: half cycle, 0%: 1 cycle, 70%: 25/30 cycle, 0%: 250/300 cycle
	Australia/New Zealand: AS/NZS CISPR:2004 Canada ICES / NMB-001:2004	
	The product was tested in a typical configuration with Agilent Technologies test systems. <sup>[1]</sup> Group 1 Class A: LC module with LAN-Interface attached.	
Safety	IEC 61010-1:2001 / EN 61010-1:2001 Canada: G1315C/65C, G1315D/G1365D: CSA C22.2 No.61010.1-04; G1315B/65B: CSA C22.2 No.1010.1:1992 + A2: 1997 USA: G1315C/65C, G1315D/65D: UL 61010-1 (2 <sup>nd</sup> Edition); G1315B/65B: UL61010A-1	

**Supplementary Information:**

This ISM device complies with Canadian ICES-001. Cet appareil ISM est conforme à la norme NMB-001 du Canada.

**This DoC applies to above-listed products placed on the EU market after:**

May 1, 2010  
Date

  
Wolfgang Wilde, Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor,  
or Agilent Technologies Deutschland GmbH, Herrenberger Straße 130, 71034 Böblingen, Germany.

Revision: D

Document No. G1315-90510





**Agilent Technologies**

## **Declaration of Conformity** to manufacturing specifications

We herewith inform you that the product  
G1316A with serial number:

**DEACN08219**

has successfully passed all our production  
quality tests.

During final instrument performance verification  
the following functional characteristics were  
individually tested for conformance with our  
internal specifications:

**Leak Sensor Function**



**Heating Performance**



**Cooling Performance**



**Operational Test**



**IEC 61010-1 Annex F Safety Tests**



September 22, 2011

Signature: 

Test Technician: Erdem Batu

**G1316A**



**DEACN08219**





**Agilent Technologies**

**DECLARATION OF CONFORMITY**

According to ISO/IEC Guide 22 and CEN/CENELEC EN 45014



**Manufacturer's Name:** Agilent Technologies Singapore (International) Pte. Ltd.  
**Manufacturer's Address:** No. 1 Yishun Ave 7  
**Supplier's Address:** SINGAPORE 768923  
Singapore

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** Thermostatted Column Compartment  
**Model Number:** G1316A, G1316B, G1316C  
**Product Options:** This declaration covers all options of the above product

**complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:**

Low Voltage Directive 2006/95/EC and  
EMC Directive 2004/108/EC

**and conforms with the following product standards:**

EMC	Standard	Limit
	IEC 61326-1 :2005 / EN 61326-1:2006 CISPR 11 / EN 55011 IEC 61000-3-2 / EN61000-3-2 IEC 61000-3-3 / EN61000-3-3 IEC 61000-4-2 / EN 61000-4-2 IEC 61000-4-3 / EN 61000-4-3 IEC 61000-4-4 / EN 61000-4-4 IEC 61000-4-5 / EN 61000-4-5 IEC 61000-4-6 / EN 61000-4-6 IEC 61000-4-11 / EN 61000-4-11	Group 1 Class B  4kV CD, 4kV AD 3 V/m: 80-1000 MHz, 1.4-2GHz; 1V/m: 2-2.7GHz 0.5 kV signal lines, 1 kV power lines 0.5 kV line-line, 1 kV line-ground 3Vrms, 0.15-80 MHz 0%: half cycle, 0%: 1 cycle, 70%: 25/30 cycle, 0%: 250/300 cycle
	Australia/New Zealand: AS/NZS CISPR:2004 Canada ICES / NMB-001:2004	
	The product was tested in a typical configuration with Agilent Technologies test systems.	
Safety	IEC 61010-1:2001 / EN 61010-1:2001 Canada: CSA C22.2 No. 1010.1:1992+A2:1997 USA: UL61010A-1	

**Supplementary Information:**

This ISM device complies with Canadian ICES-001. Cet appareil ISM est conforme à la norme NMB-001 du Canada.

**This DoC applies to above-listed products placed on the EU market after:**

May 1, 2010

Date

Wolfgang Wilde, Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor,  
or Agilent Technologies Deutschland GmbH, Herrenberger Straße 130, 71034 Böblingen, Germany.

Revision: D

Document No. G1316-90510







**Agilent Technologies**

## **Declaration of Conformity** to manufacturing specifications

We herewith inform you that the product  
G1311B with serial number:

**DEAB702998**

has successfully passed all our production  
quality tests.

During final instrument performance verification  
the following functional characteristics were  
individually tested for conformance with our  
internal specifications:

**Leak Sensor Function**



**Leak Test**



**Gradient Precision Test**



**Operational Test**



**IEC 61010-1 Annex F Safety Tests**



September 16, 2011

Signature: \_\_\_\_\_

Test Technician: Fabian Muckenfuss

**G1311B**



**DEAB702998**





**Agilent Technologies**

**DECLARATION OF CONFORMITY**  
According to ISO/IEC Guide 22 and CEN/CENELEC EN 45014



**Manufacturer's Name:** Agilent Technologies Singapore (International) Pte. Ltd.  
**Manufacturer's Address:** No. 1 Yishun Ave 7  
**Supplier's Address:** SINGAPORE 768923  
Singapore

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** Isocratic Pump, Quaternary Pump

**Model Number:** G1310A, G1310B, G1311A, G1311B, G1311C

**Product Options:** This declaration covers all options of the above products

**complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:**

Low Voltage Directive 2006/95/EC  
EMC Directive 2004/108/EC

**and conforms with the following product standards:**

EMC	Standard	Limit
	IEC 61326-1 :2005 / EN 61326-1:2006 CISPR 11 / EN 55011 IEC 61000-3-2 / EN61000-3-2 IEC 61000-3-3 / EN61000-3-3 IEC 61000-4-2 / EN 61000-4-2 IEC 61000-4-3 / EN 61000-4-3 IEC 61000-4-4 / EN 61000-4-4 IEC 61000-4-5 / EN 61000-4-5 IEC 61000-4-6 / EN 61000-4-6 IEC 61000-4-11 / EN 61000-4-11  Australia/New Zealand: AS/NZS CISPR:2004 Canada ICES / NMB-001:2004  The product was tested in a typical configuration with Agilent Technologies test systems. <sup>[1]</sup> Group 1 Class A: LC module with LAN-Interface attached.	Group 1 Class B <sup>[1]</sup>  4kV CD, 4kV AD 3 V/m: 80-1000 MHz, 1.4-2GHz; 1V/m: 2-2.7GHz 0.5 kV signal lines, 1 kV power lines 0.5 kV line-line, 1 kV line-ground 3Vrms, 0.15-80 MHz 0%: half cycle, 0%: 1 cycle, 70%: 25/30 cycle, 0%: 250/300 cycle
Safety	IEC 61010-1:2001 / EN 61010-1:2001 Canada: CSA C22.2 No. 1010.1:1992+A2:1997 USA: UL61010A-1	

**Supplementary Information:**

This ISM device complies with Canadian ICES-001. Cet appareil ISM est conforme à la norme NMB-001 du Canada.

**This DoC applies to above-listed products placed on the EU market after:**

May 1, 2010

Date

Wolfgang Wilde, Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor,  
or Agilent Technologies Deutschland GmbH, Herrenberger Straße 130, 71034 Böblingen, Germany.

Revision: C

Document No. G1311-90510





**Agilent Technologies**

## **Declaration of Conformity** to manufacturing specifications

We herewith inform you that the product  
G1329B with serial number:

**DEAB307235**

has successfully passed all our production  
quality tests.

During final instrument performance verification  
the following functional characteristics were  
individually tested for conformance with our  
internal specifications:

<b>Leak Sensor Function</b>	<input checked="" type="checkbox"/>
<b>Leak Test</b>	<input checked="" type="checkbox"/>
<b>Reproducibility Test</b>	<input checked="" type="checkbox"/>
<b>Operational Test</b>	<input checked="" type="checkbox"/>
<b>IEC 61010-1 Annex F Safety Tests</b>	<input checked="" type="checkbox"/>

September 19, 2011

Signature:   
Test Technician: Udo Ehrmann

**G1329B**



**DEAB307235**



 <b>Agilent Technologies</b>	<b>DECLARATION OF CONFORMITY</b> According to ISO/IEC Guide 22 and CEN/CENELEC EN 45014	
---	--	---

**Manufacturer's Name:** Agilent Technologies Singapore (International) Pte. Ltd.  
**Manufacturer's Address:** No. 1 Yishun Ave 7  
**Supplier's Address:** SINGAPORE 768923  
 Singapore

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** Autosampler  
**Model Number:** G1329A, G1329B, G2260A, G4303A, G5629A  
**Product Options:** This declaration covers all options of the above products

**complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:**

Low Voltage Directive 2006/95/EC  
 EMC Directive 2004/108/EC

**and conforms with the following product standards:**


EMC	Standard	Limit
	IEC 61326-1 :2005 / EN 61326-1:2006 CISPR 11 / EN 55011 IEC 61000-3-2 / EN61000-3-2 IEC 61000-3-3 / EN61000-3-3 IEC 61000-4-2 / EN 61000-4-2 IEC 61000-4-3 / EN 61000-4-3 IEC 61000-4-4 / EN 61000-4-4 IEC 61000-4-5 / EN 61000-4-5 IEC 61000-4-6 / EN 61000-4-6 IEC 61000-4-11 / EN 61000-4-11	Group 1 Class B <sup>[1]</sup>  4kV CD, 4kV AD 3 V/m: 80-1000 MHz, 1.4-2GHz; 1V/m: 2-2.7GHz 0.5 kV signal lines, 1 kV power lines 0.5 kV line-line, 1 kV line-ground 3Vrms, 0.15-80 MHz 0%: half cycle, 0%: 1 cycle, 70%: 25/30 cycle, 0%: 250/300 cycle
	Australia/New Zealand: AS/NZS CISPR:2004 Canada ICES / NMB-001:2004	
	The product was tested in a typical configuration with Agilent Technologies test systems. <sup>[1]</sup> Group 1 Class A: LC module with LAN-Interface attached.	
<b>Safety</b>	IEC 61010-1:2001 / EN 61010-1:2001 Canada: CSA C22.2 No. 1010.1:1992+A2:1997 USA: UL61010A-1	

**Supplementary Information:**

This ISM device complies with Canadian ICES-001. Cet appareil ISM est conforme à la norme NMB-001 du Canada.

**This DoC applies to above-listed products placed on the EU market after:**

May 1, 2010  
 Date

  
 Wolfgang Wilde, Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor,  
 or Agilent Technologies Deutschland GmbH, Herrenberger Straße 130, 71034 Böblingen, Germany.

Revision: D

Document No. G1329-90510

