

# List Certificate

Dual temperature probes, part number 02 101581 00



## VARIAN

Varian Australia Pty Ltd  
679 Springvale Road  
Mulgrave VIC 3170  
Australia

|                                  |                      |
|----------------------------------|----------------------|
| Operator: <u>Van D. T. T.</u>    | Date: <u>13.9.10</u> |
| Serial number: <u>UV1009M382</u> |                      |

### Temperature accuracy

|         | Reference temperature |                  | Probe temperature |    | Customer specification |
|---------|-----------------------|------------------|-------------------|----|------------------------|
|         | Target                | Actual           |                   | °C |                        |
| Probe 1 | 0.0                   | <u>-0.001</u> °C | <u>-0.01</u>      | °C | ± 0.3 °C               |
|         | 100.0                 | <u>86.38</u> °C  | <u>86.37</u>      | °C | ± 0.3 °C               |
|         | 50.0                  | <u>49.09</u> °C  | <u>49.08</u>      | °C | ± 0.3 °C               |
| Probe 2 | 0.0                   | <u>-0.001</u> °C | <u>0.00</u>       | °C | ± 0.3 °C               |
|         | 100.0                 | <u>86.38</u> °C  | <u>86.37</u>      | °C | ± 0.3 °C               |
|         | 50.0                  | <u>49.09</u> °C  | <u>49.08</u>      | °C | ± 0.3 °C               |